



## FR4003 Firmware history 2.22

Ver.	Date	Applied Modifications
2.22	31/07/2019	<ul style="list-style-type: none"><li>– Implemented fast upgrade</li><li>– Minor bug fix</li><li>– Minor improvements</li></ul>
2.21	21/06/2019	<ul style="list-style-type: none"><li>– Improved UART communication</li><li>– Added the frequency of 15 MHz in the autocal</li></ul>
2.09	21/03/2019	<ul style="list-style-type: none"><li>– Improved weighing of AVG detector</li></ul>
2.05	06/12/2017	<ul style="list-style-type: none"><li>– Minor bugs fixed</li></ul>
2.03	18/09/2017	<ul style="list-style-type: none"><li>– Activated FFT</li><li>– Mil filters up to 10 Hz</li></ul>
1.22	25/08/2016	<ul style="list-style-type: none"><li>– Auto switch off when battery is &lt; 6.5 V</li></ul>
1.21	18/03/2016	<ul style="list-style-type: none"><li>– Added possibility of connection to 9010F</li></ul>
1.17	02/12/2015	<ul style="list-style-type: none"><li>– Improved value levels during autocalibration</li></ul>
1.15	17/11/2015	<ul style="list-style-type: none"><li>– Modified, internal generator's level, during autocalibration</li></ul>
1.13	26/07/2015	<ul style="list-style-type: none"><li>– First release</li></ul>